

(PCT)

ИммуноТД (Иммунотурбидиметрический)



Параметры для ввода в программу анализатора Beckman-Coulter AU 680

Parameters		Specific Test Parameters			
General	LIH	ISE	HbA1c	Calculated Test	Range
Test Name: <input type="text" value="PCT"/> < > Type: <input type="text" value="Serum"/> Operation <input type="text" value="Yes"/>					
Sample Volume <input type="text" value="8.8"/> μL Dilution <input type="text" value="0"/> μL OD Limit <input type="text" value="Min.OD -2.0"/> <input type="text" value="Max.OD 2.5"/>					
Pre-Dilution Rate <input type="text" value="1"/> ∇ Rgt. Volume R1(R1-1) <input type="text" value="105"/> μL Dilution <input type="text" value="0"/> μL Reagent OD Limit <input type="text" value="First Low -2.0"/> <input type="text" value="High 2.5"/>					
R2(R2-1) <input type="text" value="35"/> μL Dilution <input type="text" value="0"/> μL Last Low <input type="text" value="-2.0"/> <input type="text" value="High 2.5"/>					
Wavelength Pri <input type="text" value="660"/> ∇nm Sec. <input type="text" value="NONE"/> ∇nm Dynamic Range Low <input type="text" value="0"/> <input type="text" value="High 35"/>					
Method <input type="text" value="RATE"/> ∇ Correlation Factor A <input type="text" value="1"/> <input type="text" value="B 0"/>					
Reaction Slope <input type="text" value="+"/> ∇ Factor for Maker A <input type="text" value="1"/> <input type="text" value="B 0"/>					
Measuring Point1 First <input type="text" value="14"/> Last <input type="text" value="23"/>					
Measuring Point2 First <input type="text" value=""/> Last <input type="text" value=""/>					
Linearity Limit <input type="text" value=""/> %					
Lag Time Check <input type="text" value=""/>					
Onboard Stability Period <input type="text" value="999"/> Day <input type="text" value=""/> Hour					

Parameters		Specific Test Parameters			
General	LIH	ISE	HbA1c	Calculated Test	Range
Test Name: <input type="text" value="PCT"/> < > Type: <input type="text" value="Serum"/>					
Value/Flag: <input type="text" value=""/> Level L: <input type="text" value=""/> Level H: <input type="text" value=""/>					
Specific Ranges:					
From To Low High					
Sex Year Month Year Month					
1. <input type="text" value=""/> <input type="text" value=""/> <input type="text" value=""/> <input type="text" value=""/> <input type="text" value=""/> <input type="text" value=""/>					
2. <input type="text" value=""/> <input type="text" value=""/> <input type="text" value=""/> <input type="text" value=""/> <input type="text" value=""/> <input type="text" value=""/>					
3. <input type="text" value=""/> <input type="text" value=""/> <input type="text" value=""/> <input type="text" value=""/> <input type="text" value=""/> <input type="text" value=""/>					
4. <input type="text" value=""/> <input type="text" value=""/> <input type="text" value=""/> <input type="text" value=""/> <input type="text" value=""/> <input type="text" value=""/>					
5. <input type="text" value=""/> <input type="text" value=""/> <input type="text" value=""/> <input type="text" value=""/> <input type="text" value=""/> <input type="text" value=""/>					
6. <input type="text" value=""/> <input type="text" value=""/> <input type="text" value=""/> <input type="text" value=""/> <input type="text" value=""/> <input type="text" value=""/>					
7. No demographics <input type="text" value="0"/> <input type="text" value="0.5"/>					
8. Not within expected values <input type="text" value=""/> <input type="text" value=""/>					
Unit <input type="text" value="ng/ml"/> Decimal Places <input type="text" value="2"/>					
Panic Value <input type="text" value="Low"/> <input type="text" value="High"/>					

Parameters		Calibration Parameters			
Calibrators	Calibration Specific	STAT Table Calibration			
General	ISE				
Test Name: <input type="text" value="PCT"/> < > Type: <input type="text" value="Serum"/> <input type="radio"/> Use Serum Cal.					
Calibration Type: <input type="text" value="6AB"/> ∇ Formula: <input type="text" value="Spline"/> ∇ Counts: <input type="text" value="2"/> ∇					
<Calibrator Parameters>					
Calibrator OD Conc Low High					
Point 1: <input type="text" value="TruCal PCT 0"/> <input type="text" value=""/> <input type="text" value="0"/> <input type="text" value="-0.1"/> <input type="text" value="2.5"/>					
Point 2: <input type="text" value="TruCal PCT 1"/> <input type="text" value=""/> <input type="text" value="*"/> <input type="text" value="-0.1"/> <input type="text" value="2.5"/>					
Point 3: <input type="text" value="TruCal PCT 2"/> <input type="text" value=""/> <input type="text" value="*"/> <input type="text" value="-0.1"/> <input type="text" value="2.5"/>					
Point 4: <input type="text" value="TruCal PCT 3"/> <input type="text" value=""/> <input type="text" value="*"/> <input type="text" value="-0.1"/> <input type="text" value="2.5"/>					
Point 5: <input type="text" value="TruCal PCT 4"/> <input type="text" value=""/> <input type="text" value="*"/> <input type="text" value="-0.1"/> <input type="text" value="2.5"/>					
Point 6: <input type="text" value="TruCal PCT 5"/> <input type="text" value=""/> <input type="text" value="*"/> <input type="text" value="-0.1"/> <input type="text" value="2.5"/>					
Point 7: <input type="text" value=""/> <input type="text" value=""/> <input type="text" value=""/> <input type="text" value=""/> <input type="text" value=""/>					
Point 8: <input type="text" value=""/> <input type="text" value=""/> <input type="text" value=""/> <input type="text" value=""/> <input type="text" value=""/>					
Point 9: <input type="text" value=""/> <input type="text" value=""/> <input type="text" value=""/> <input type="text" value=""/> <input type="text" value=""/>					
Point 10: <input type="text" value=""/> <input type="text" value=""/> <input type="text" value=""/> <input type="text" value=""/> <input type="text" value=""/>					
Slope Check <input type="text" value="None"/> ∇					
Allowance Range Check					
<input type="radio"/> Reagent Blank <input type="text" value=""/>					
<input type="radio"/> Calibration <input type="text" value=""/>					
Advanced Calibration					
Operation <input type="text" value=""/> ∇					
Interval (RB/ACAL) <input type="text" value=""/> ∇					
<Point Cal. For No. of Correction Points <input type="text" value=""/> ∇ Use Master Curve <input type="text" value=""/> ∇ <input type="radio"/> Lot Calibration					
Master Curve>					
Calibrator OD Conc Low High Stability					
Point 1: <input type="text" value=""/> <input type="text" value=""/> <input type="text" value=""/> <input type="text" value=""/> <input type="text" value=""/>					
Point 2: <input type="text" value=""/> <input type="text" value=""/> <input type="text" value=""/> <input type="text" value=""/> <input type="text" value=""/>					
MB Type Factor: <input type="text" value=""/> 1-Point Calibration Point <input type="text" value=""/> ∇ <input type="radio"/> with Conc-0					
Reagent Blank <input type="text" value=""/> Day <input type="text" value=""/> Hour					
Calibration <input type="text" value=""/> Day <input type="text" value=""/> Hour					

Диапазон нормальных значений указан в соответствии с рекомендациями производителя.

При использовании единиц измерения, отличающихся от приведенных, убедитесь, что значения стандартов, контрольных материалов, диапазонов нормальных значений и линейности метода введены в этих же единицах.

* -вводится из паспорта к калибратору (TruCal PCT, включает калибратор с конц. 0).

Контроль по TruLab PCT уровень 1 и уровень 2.